

Certificate of Analysis

Certificate No: BMDX/COA/P/200/2025

Product Name: Mueller Hinton II Agar

Product Code: FP90M2001

Lot No: 251113001

Expiry date: 19/02/2026

Date of Analysis: 19/11/2025 – 20/11/2025

Parameter Requirement: Physical, Chemical and Microbiological Analysis

Physical & Chemical Test	Specification	Results	
Color	Light to medium, yellow to tan, trace hazy to slightly hazy	Passed	
pH	7.30 ± 0.1	Passed	
Hardness	Semi-solid with a firm, smooth, & bounces surface	Passed	
Pour Depth/Thickness	4.00-4.50mm	Passed	
Microbiological Test	Antimicrobial Agent	Acceptable range (mm or (µg/mL)	Zone Diameter (mm) or MIC (µg/mL)
<i>Escherichia coli</i> ATCC® 25922	Gentamicin 10 µg Amikacin 30 µg Ampicillin 10 µg Ceftazidime 30 µg Tigecycline 15 µg	19-26 19-26 15-22 25-32 20-27	26 24 20 30 26
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Ceftazidime 30 µg Amikacin 30 µg Cefepime 30 µg Imipenem 10 µg Ciprofloxacin 5 µg Gentamicin 10 µg Colistin	22-29 20-26 25-31 20-28 25-33 17-23 0.25-2	23 26 29 24 30 22 2.0
<i>Enterococcus faecalis</i> ATCC® 29212	Ampicillin 10 µg Trimethoprim-sulfamethoxazole 1.25/23.75 µg Nitrofurantoin 300 µg Penicillin	15-21 ≥ 20 18-24 1-4	20 29 23 3.0
<i>Staphylococcus aureus</i> ATCC® 25923	Penicillin 10 units Gentamicin 10 µg Erythromycin 15 µg Clindamycin 2 µg Trimethoprim-sulfamethoxazole 1.25/23.75 µg	26-37 19-27 22-30 24-30 24-32	37 27 28 28 31
<i>Staphylococcus aureus</i> ATCC® 29213	Vancomycin	0.5-2	0.50
Uninoculated plate	No growth	-	-

Test Statement:

1. Incubation at $35 \pm 2^{\circ}\text{C}$ for 24 hours in ambient air.
2. Antibiotic susceptibility tests are performed in accordance with, and meet the acceptance limits of, the current ISO/TS 16782. Performance is assessed using CLSI methodology.

Sample statement:

1. Culture media were prepared according to procedure SP-PD-BMDX-02 Production Process Control and SP-PD-BMDX-11 Product Specification.
2. Culture media were kept at temperature conditions $2\text{--}8^{\circ}\text{C}$.
3. QC Organisms were prepared and followed according to procedure SP-QC-BMDX-07 Quality Control Testing and SP-PD-BMDX-11 Product Specification.

Result:

The information given is believed to be correct, all results reported in this certificate relate only to the product in this certificate of analysis according to SP-PD-BMDX-11 Product Specification.

Sign by:



Fairuz Ayuni Binti Abdul Halim

QC Executive

BSc. in Microbiology with Honours

Universiti Putra Malaysia



Biomed MDX Sdn Bhd
8, Jalan IAN 3, Industri Angkasa Nuri,
76100 Durian Tunggal, Melaka, Malaysia

Date: 20/11/2025

