

Certificate of Analysis

Certificate No: BMDX/COA/P/346/2026

Product Name: Nutrient Agar
Product Code: FP90N1002
Lot No: 260305001
Expiry date: 11/06/2026
Date of Analysis: 05/03/2026 – 07/03/2026
Parameter Requirement: Physical, Chemical and Microbiological Analysis

| Physical & Chemical Test | Specification | Results |
|--------------------------|---|---------|
| Color | Straw colored-gel | Passed |
| pH | 7.40 ± 0.2 | Passed |
| Hardness | Semi-solid with a firm, smooth, & bounces surface | Passed |
| Pour Depth/Thickness | 3.00-3.50mm | Passed |

| Microbiological Test | Specification | Results |
|---|---------------|---------|
| <i>Enterococcus faecalis</i> ATCC® 29212 | Growth | Passed |
| <i>Pseudomonas aeruginosa</i> ATCC® 27853 | Growth | Passed |
| <i>Escherichia coli</i> ATCC® 25922 | Growth | Passed |
| Uninoculated plate | No Growth | Passed |

Sample statement:

1. Culture media were prepared according to procedure SP-PD-BMDX-02 Production Process Control and SP-PD-BMDX-11 Product Specification.
2. Culture media were kept at temperature conditions 2-8°C.
3. QC Organisms were prepared and followed according to procedure SP-QC-BMDX-07 Quality Control Testing and SP-PD-BMDX-11 Product Specification.

Result:

The information given is believed to be correct, all results reported in this certificate relate only to the product in this certificate of analysis according to SP-PD-BMDX-11 Product Specification.

Sign by:



Fairuz Ayuni Binti Abdul Halim
QC Executive
BSc. in Microbiology with Honours
Universiti Putra Malaysia

Date: 09/03/2026