

## Certificate of Analysis

Certificate No: BMDX/COA/P/430/2026

**Product Name:** Sabouraud Dextrose Agar

**Product Code:** FP90S1001

**Lot No:** 260507001

**Expiry date:** 13/08/2026

**Date of Analysis:** 08/05/2026– 11/05/2026

**Parameter Requirement:** Physical, Chemical and Microbiological Analysis

Physical & Chemical Test	Specification	Results
Color	Pale to medium, yellow to tan, clear to slightly hazy	Passed
pH	5.6 ± 0.2	Passed
Hardness	Semi-solid with a firm, smooth, & bounces surface	Passed
Pour Depth/Thickness	≥ 5.00mm	Passed

  

Microbiological Test	Specification	Results
Candida albicans ATCC® 60193	Growth at 72 hours	Passed
Trichophyton mentagrophytes ATCC® 9533	Growth at 72 hours	Passed
Uninoculated plate	No growth	Passed

### Sample statement:

1. Culture media were prepared according to procedure SP-PD-BMDX-02 Production Process Control and SP-PD-BMDX-11 Product Specification.
2. Culture media were kept at temperature conditions 2-8°C.
3. QC Organisms were prepared and followed according to procedure SP-QC-BMDX-07 Quality Control Testing and SP-PD-BMDX-11 Product Specification.

### Result:

The information given is believed to be correct, all results reported in this certificate relate only to the product in this certificate of analysis according to SP-PD-BMDX-11 Product Specification.

### Sign by:



**Fairuz Ayuni Binti Abdul Halim**

**QC Executive**

BSc. in Microbiology with Honours

Universiti Putra Malaysia

**Date:** 12/05/2026

