

Certificate of Analysis

Certificate No: BMDX/COA/P/155/2025

Product Name: MacConkey III Agar

Product Code: FP90M1002

Lot No: 251014002

Expiry Date: 19/01/2026

Date of Analysis: 15/10/2025 – 16/10/2025

Parameter Requirement: Physical, Chemical and Microbiological Analysis

Physical & Chemical Test	Specification	Results
Color	Dark red colored gel	Passed
pH	7.10 ± 0.2	Passed
Hardness	Semi-solid with a firm, smooth, & bounces surface	Passed
Pour Depth/Thickness	3.00-3.50mm	Passed

Microbiological Test	Specification	Results
<i>Escherichia coli</i> ATCC® 25922	Pink to red growth	Passed
<i>Proteus mirabilis</i> ATCC® 12453	Growth. Colorless Inhibition of swarming	Passed
<i>Salmonella choleraesuis</i> subsp. <i>choleraesuis</i> serotype	Growth. Colorless	Passed
<i>Typhimurium</i> ATCC® 14028		
<i>Staphylococcus aureus</i> ATCC® 25923	No growth	Passed
<i>Enterococcus faecalis</i> ATCC® 29212	No growth	Passed
Uninoculated plate	No growth	Passed

Sample statement:

1. Culture media were prepared according to procedure SP-PD-BMDX-02 Production Process Control and SP-PD-BMDX-11 Product Specification.
2. Culture media were kept at temperature conditions 2-8°C.
3. QC Organisms were prepared and followed according to procedure SP-QC-BMDX-07 Quality Control Testing and SP-PD-BMDX-11 Product Specification.

Result:

The information given is believed to be correct, all results reported in this certificate relate only to the product in this certificate of analysis according to SP-PD-BMDX-11 Product Specification.

Sign by:



Fairuz Ayuni Binti Abdul Halim

QC Executive

BSc. in Microbiology with Honours
Universiti Putra Malaysia

Date: 16/10/2025



Biomed MDX Sdn Bhd
8, Jalan IAN 3, Industri Angkasa Nuri,
76100 Durian Tunggal, Melaka, Malaysia