

Certificate of Analysis

Certificate No: BMDX/COA/P/247/2025

Product Name: Columbia Agar with 5% Sheep Blood / SDA Agar

Product Code: SP90C1021

Lot No: 251211002

Expired Date: 12/03/2026

Date of Analysis: 11/12/2025 – 12/12/2025

Parameter Requirement: Physical, Chemical and Microbiological Analysis

Physical & Chemical Test	Specification	Results
Color	Columbia Agar with 5% Sheep Blood: Cherry red, opaque, no hemolysis	Passed
	SDA: Pale to medium, yellow to tan, clear to slightly hazy.	Passed
pH	Columbia Agar with 5% Sheep Blood: 7.3 ± 0.2	Passed
	SDA: 5.6 ± 0.2	Passed
Hardness	Semi-solid with a firm, smooth, & bounces surface	Passed
Pour Depth/Thickness	Columbia Agar with 5% Sheep Blood: 3.00mm to 3.50mm	Passed
	SDA: 3.00mm to 3.50mm	Passed

Microbiological Test	Specification	Results
Columbia Agar with 5% Sheep Blood:		
<i>E. coli</i> ATCC® 25922	Growth at 24 hours, beta hemolysis	Passed
<i>E. faecalis</i> ATCC® 9533	Growth at 24 hours, gamma hemolysis	Passed
<i>S. pyogenes</i> ATCC® 19615	Growth at 24 hours, beta hemolysis	Passed
<i>S. pneumoniae</i> ATCC® 6305	Growth at 24 hours, alpha hemolysis	Passed
<i>C. albicans</i> ATCC® 60193	Growth at 24 hours, no hemolysis	Passed
Uninoculated plate	No growth	Passed
SDA:		
<i>C. albicans</i> ATCC® 60193	Growth. Straw colonies	Passed
<i>T. mentagrophytes</i> ATCC® 9533	Growth. Pale straw-colored colonies.	Passed
Uninoculated plate	No growth	Passed

Sample statement:

1. Culture media were prepared according to procedure SP-PD-BMDX-02 Production Process Control and SP-PD-BMDX-11 Product Specification.
2. Culture media were kept at temperature conditions 2-8°C.
3. QC Organisms were prepared and followed according to procedure SP-QC-BMDX-07 Quality Control Testing and SP-PD-BMDX-11 Product Specification.



Result:

The information given is believed to be correct, all results reported in this certificate relate only to the product in this certificate of analysis according to SP-PD-BMDX-11 Product Specification.

Sign by:

Fairuz Ayuni Binti Abdul Halim

QC Executive

BSc. in Microbiology with Honours

Universiti Putra Malaysia

Date: 15/12/2025



Biomed MDX Sdn Bhd
8, Jalan IAN 3, Industri Angkasa Nuri,
76100 Durian Tunggal, Melaka, Malaysia