

Certificate of Analysis

Certificate No: BMDX/COA/P/355/2026

Product Name: CHROMagar MRSA
Product Code: FP90M4004
Lot No: 260310002
Expiry date: 16/06/2026
Date of Analysis: 11/03/2026 – 13/03/2026
Parameter Requirement: Physical, Chemical and Microbiological Analysis

Physical & Chemical Test	Specification	Results
Color	Light straw	Passed
pH	6.9 ± 0.2	Passed
Hardness	Semi-solid with a firm, smooth, & bounces surface	Passed
Pour Depth/Thickness	4.00-4.50mm	Passed

Microbiological Test	Specification	Results
<i>MR S. aureus ATCC® 33592</i>	Growth. Mauve	Passed
<i>Staphylococcus aureus ATCC® 25923</i>	No growth	Passed
<i>Enterococcus faecalis ATCC® 29212</i>	No growth	Passed
<i>Escherichia coli ATCC® 25922</i>	No growth	Passed
<i>Candida albicans ATCC® 10231</i>	No growth	Passed
Uninoculated plate	No growth	Passed

Sample statement:

1. Culture media were prepared according to procedure SP-PD-BMDX-02 Production Process Control and SP-PD-BMDX-11 Product Specification.
2. Culture media were kept at temperature conditions 2-8°C.
3. QC Organisms were prepared and followed according to procedure SP-QC-BMDX-07 Quality Control Testing and SP-PD-BMDX-11 Product Specification.

Result:

The information given is believed to be correct, all results reported in this certificate relate only to the product in this certificate of analysis according to SP-PD-BMDX-11 Product Specification.

Sign by:



Fairuz Ayuni Binti Abdul Halim

QC Executive

BSc. in Microbiology with Honours

Universiti Putra Malaysia

Date: 13/03/2026

