

Certificate of Analysis

Certificate No: BMDX/COA/P/402/2026

Product Name: Xylose Lysine Deoxycholate Agar
Product Code: FP90X1002
Lot No: 260416002
Expiry date: 11/06/2026
Date of Analysis: 17/04/2026 – 21/04/2026
Parameter Requirement: Physical, Chemical and Microbiological Analysis

Physical & Chemical Test	Specification	Results
Color	Red colored gel	Passed
pH	7.40 ± 0.2	Passed
Hardness	Semi-solid with a firm, smooth, & bounces surface	Passed
Pour Depth/Thickness	3.00-3.50mm	Passed

Microbiological Test	Specification	Results
<i>Shigella flexneri</i> ATCC® 12022	Irregular red colonies	Passed
<i>Salmonella enterica</i> subsp. <i>enterica</i> serotype <i>Typhimurium</i> ATCC® 14028	Red colonies, black centers	Passed
<i>Escherichia coli</i> ATCC® 25922	No growth or yellow colonies	Passed
<i>Enterococcus faecalis</i> ATCC® 29212	Inhibition partial to complete	Passed
Uninoculated plate	No growth	Passed

Sample statement:

- Culture media were prepared according to procedure SP-PD-BMDX-02 Production Process Control and SP-PD-BMDX-11 Product Specification.
- Culture media were kept at temperature conditions 2-8°C.
- QC Organisms were prepared and followed according to procedure SP-QC-BMDX-07 Quality Control Testing and SP-PD-BMDX-11 Product Specification.

Result:

The information given is believed to be correct, all results reported in this certificate relate only to the product in this certificate of analysis according to SP-PD-BMDX-11 Product Specification.

Sign by:



Fairuz Ayuni Binti Abdul Halim

QC Executive

BSc. in Microbiology with Honours

Date: 22/04/2026

