

## Certificate of Analysis

Certificate No: BMDX/COA/P/356/2026

**Product Name:** CHROMagar™ Candida Plus Agar

**Product Code:** FP90C5004

**Lot No:** 260310003

**Expiry date:** 16/06/2026

**Date of Analysis:** 11/03/2026 – 13/03/2026

**Parameter Requirement:** Physical, Chemical and Microbiological Analysis

Physical & Chemical Test	Specification	Results
Color	Light straw	Passed
pH	6.1 ± 0.2	Passed
Hardness	Semi-solid with a firm, smooth & bounces surface	Passed
Pour Depth/Thickness	4.00-4.50mm	Passed

Microbiological Test	Specification	Results
<i>Candida albicans</i> ATCC® 60193	Growth; Green-blue	Passed
<i>Candida tropicalis</i> ATCC® 1369	Growth; Metallic blue with pink halo	Passed
<i>Candida krusei</i> ATCC® 14243	Growth; Pink and fuzzy	Passed
<i>Candida glabrata</i> ATCC® 2001	Growth; Mauve	Passed
<i>Escherichia coli</i> ATCC® 25922	No growth	Passed
Uninoculated plate	No growth	Passed

### Sample statement:

1. Culture media were prepared according to procedure SP-PD-BMDX-02 Production Process Control and SP-PD-BMDX-11 Product Specification.
2. Culture media were kept at temperature conditions 2-8°C.
3. QC Organisms were prepared and followed according to procedure SP-QC-BMDX-07 Quality Control Testing and SP-PD-BMDX-11 Product Specification.

### Result:

The information given is believed to be correct, all results reported in this certificate relate only to the product in this certificate of analysis according to SP-PD-BMDX-11 Product Specification.

Sign by:



Fairuz Ayuni Binti Abdul Halim

QC Executive

BSc. in Microbiology with Honours

Universiti Putra Malaysia

Date: 13/03/2026

