

Certificate of Analysis

Certificate No: BMDX/COA/P/441/2026

Product Name: Nutrient Agar
Product Code: FP90N1002
Lot No: 260513003
Expiry date: 19/08/2026
Date of Analysis: 13/05/2026 – 15/05/2026
Parameter Requirement: Physical, Chemical and Microbiological Analysis

Physical & Chemical Test	Specification	Results
Color	Straw colored-gel	Passed
pH	7.40 ± 0.2	Passed
Hardness	Semi-solid with a firm, smooth, & bounces surface	Passed
Pour Depth/Thickness	3.00-3.50mm	Passed

Microbiological Test	Specification	Results
<i>Enterococcus faecalis</i> ATCC® 29212	Growth	Passed
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Growth	Passed
<i>Escherichia coli</i> ATCC® 25922 Uninoculated plate	Growth No Growth	Passed Passed

Sample statement:

1. Culture media were prepared according to procedure SP-PD-BMDX-02 Production Process Control and SP-PD-BMDX-11 Product Specification.
2. Culture media were kept at temperature conditions 2-8°C.
3. QC Organisms were prepared and followed according to procedure SP-QC-BMDX-07 Quality Control Testing and SP-PD-BMDX-11 Product Specification.

Result:

The information given is believed to be correct, all results reported in this certificate relate only to the product in this certificate of analysis according to SP-PD-BMDX-11 Product Specification.

Sign by:



Fairuz Ayuni Binti Abdul Halim
QC Executive
BSc. in Microbiology with Honours
Universiti Putra Malaysia

Date: 15/05/2026

