

## Certificate of Analysis

Certificate No: BMDX/COA/P/357/2026

**Product Name:** Schaedler Agar with Vitamin K1 And 5% Sheep Blood Agar

**Product Code:** FP90S2001

**Lot No:** 260310004

**Expiry date:** 02/06/2026

**Date of Analysis:** 11/03/2026 – 13/03/2026

**Parameter Requirement:** Physical and Microbiological Analysis

Physical & Chemical Test	Specification	Results
Color	Red to dark red	Passed
pH	7.6 ± 0.2	Passed
Depth	3.00mm-3.50mm	Passed
Physical Appearance/Hardness	Semi-solid with a firm, smooth, & bounces surface.	Passed

Microbiological Test	Specification	Results
<i>Bacteroides fragilis</i> ATCC® 25285	Growth	Passed
<i>Clostridium perfringens</i> ATCC® 13124	Growth	Passed
Uninoculated plate	No Growth	Passed

### Sample statement:

- Culture media were prepared according to procedure per SP-PD-BMDX-02 Production Process Control and SP-PD-BMDX-11 Product Specification.
- Culture media were kept at temperature conditions 2-8°C.
- QC Organisms were prepared and followed according to procedure SP-QC-BMDX-07 Quality Control Testing and SP-PD-BMDX-11 Product Specification.

### Result:

The information given is believed to be correct, all results reported in this certificate relate only to the product in this certificate of analysis according to SP-PD-BMDX-11 Product Specification.

### Sign by:



Fairuz Ayuni Binti Abdul Halim

QC Executive

BSc. in Microbiology with Honours

Universiti Putra Malaysia

Date: 13/03/2026

