

Certificate of Analysis

Certificate No: BMDX/COA/P/218/2025

Product Name: Columbia Agar with 5% Sheep Blood / MacConkey III

Product Code: SP90C2022

Lot No: 251125003

Expiry date: 24/02/2026

Date of Analysis: 27/11/2025 – 28/11/2025


Parameter Requirement: Physical, Chemical and Microbiological Analysis

Physical & Chemical Test	Specification	Results
Color	Columbia Agar with 5% Sheep Blood: Red	Passed
	MacConkey III Agar: Dark red colored gel	Passed
	Columbia Agar with 5% Sheep Blood: 7.3 ± 0.2	Passed
pH	MacConkey III Agar: 7.1 ± 0.2	Passed
	Semi-solid with a firm, smooth, & bounces surface	Passed
Hardness	Columbia Agar with 5% Sheep Blood: 3.00mm to 3.50mm	Passed
Pour Depth/Thickness	MacConkey III Agar: 3.00mm to 3.50mm	Passed

Microbiological Test	Specification	Results
Columbia Agar with 5% Sheep Blood:		
<i>E. coli</i> ATCC® 25922	Growth at 24 hours, beta hemolysis	Passed
<i>E. faecalis</i> ATCC® 29212	Growth at 24 hours, gamma hemolysis	Passed
<i>S. pyogenes</i> ATCC® 19615	Growth at 24 hours, beta hemolysis	Passed
<i>S. pneumoniae</i> ATCC® 6305	Growth at 24 hours, alpha hemolysis	Passed
<i>C. albicans</i> ATCC® 60193	Growth at 24 hours, white colonies, no hemolysis	Passed
Uninoculated plate	No growth	Passed
MacConkey III Agar:		
<i>E. coli</i> ATCC® 25922	Pink to red growth	Passed
<i>P. mirabilis</i> ATCC® 12453	Growth colorless inhibition of swarming	Passed
<i>Salmonella choleraesuis</i> subsp. <i>choleraesuis</i> serotype	Growth. Colorless	Passed
<i>Typhimurium</i> ATCC® 14028	No growth	Passed
<i>E. faecalis</i> ATCC® 29212	No growth	Passed
<i>S. aureus</i> ATCC® 25923	No growth	Passed
Uninoculated plate		

Sample statement:

1. Culture media were prepared according to procedure SP-PD-BMDX-02 Production Process Control and SP-PD-BMDX-11 Product Specification.
2. Culture media were kept at temperature conditions 2-8°C.
3. QC Organisms were prepared and followed according to procedure SP-QC-BMDX-07 Quality Control Testing and SP-PD-BMDX-11 Product Specification.



Biomed MDX Sdn Bhd
8, Jalan IAN 3, Industri Angkasa Nuri,
76100 Durian Tunggal, Melaka, Malaysia

Result:

The information given is believed to be correct, all results reported in this certificate relate only to the product in this certificate of analysis according to SP-PD-BMDX-11 Product Specification.

Sign by:



Fairuz Ayuni Binti Abdul Halim

QC Executive

BSc. in Microbiology with Honours

Universiti Putra Malaysia

Date: 28/11/2025



Biomed MDX Sdn Bhd
8, Jalan IAN 3, Industri Angkasa Nuri,
76100 Durian Tunggal, Melaka, Malaysia

