

Certificate of Analysis

Certificate No: BMDX/COA/P/135/2025

Product Name: Chocolate Agar with Bacitracin
Product Code: FP90C3011
Lot No: 250929001
Expiry date: 05/01/2026
Date of Analysis: 30/09/2025 – 01/10/2025
Parameter Requirement: Physical, Chemical and Microbiological Analysis

Physical & Chemical Test	Specification	Results
Color	Dark Brown	Passed
pH	7.3 ± 0.2	Passed
Hardness	Semi-solid with a firm, smooth, & bounces surface	Passed
Pour Depth/Thickness	3.00-3.50mm	Passed

Microbiological Test	Specification	Results
<i>Haemophilus influenzae</i> ATCC® 10211	Growth	Passed
<i>Staphylococcus aureus</i> ATCC® 25923	No growth	Passed
Uninoculated plate	No growth	Passed

Sample statement:

1. Culture media were prepared according to procedure SP-PD-BMDX-02 Production Process Control and SP-PD-BMDX-11 Product Specification.
2. Culture media were kept at temperature conditions 2-8°C.
3. QC Organisms were prepared and followed according to procedure SP-QC-BMDX-07 Quality Control Testing and SP-PD-BMDX-11 Product Specification.

Result:

The information given is believed to be correct, all results reported in this certificate relate only to the product in this certificate of analysis according to SP-PD-BMDX-11 Product Specification.

Sign by:



Fairuz Ayuni Binti Abdul Halim
QC Executive
BSc. in Microbiology with Honours
Universiti Putra Malaysia

Date: 01/10/2025

