

Certificate of Analysis

Certificate No: BMDX/COA/P/039/2025

Product Name: DNase Agar

Product Code: FP90D1001

Lot No: 250530002

Date of Analysis: 03/06/2025 – 04/06/2025

Parameter Requirement: Physical, Chemical and Microbiological Analysis

Physical & Chemical Test	Specification	Results
Color	Light to medium amber, slightly opalescent with slight precipitate	Passed
pH	7.30 ± 0.2	Passed
Hardness	Semi-solid with a firm, smooth, & bounces surface	Passed
Pour Depth/Thickness	3.00-3.50mm	Passed

Microbiological Test	Specification	Results
<i>Staphylococcus aureus</i> ATCC® 25923	Growth. DNase Positive.	Passed
<i>Streptococcus pyogenes</i> ATCC® 19615	Growth. DNase Positive.	Passed
<i>Staphylococcus epidermis</i> ATCC® 12228	Growth. DNase Negative.	Passed
Uninoculated plate	No Growth	Passed

Sample statement:

1. Culture media were prepared according to procedure SP-PD-BMDX-02 Production Process Control and SP-PD-BMDX-11 Product Specification.
2. Culture media were control at temperature conditions 2-8°C.
3. QC Organisms were prepared and followed according to procedure SP-QC-BMDX-07 Quality Control Testing and SP-PD-BMDX-11 Product Specification.

Result:

The information given is believed to be correct, all results reported in this certificate relate only to the product in this certificate of analysis according to SP-PD-BMDX-11 Product Specification.

Sign by:



Fairuz Ayuni Binti Abdul Halim

QC Executive

BSc. in Microbiology with Honours

Universiti Putra Malaysia

Date: 04/06/2025



Biomed MDX Sdn Bhd
8, Jalan IAN 3, Industri Angkasa Nuri,
76100 Durian Tunggal, Melaka, Malaysia