



Certificate of Analysis

Certificate No: BMDX/COA/P/006/2025

Product Name: Mueller Hinton II Agar

Product Code: FP90M2001

Lot No: 250207001

Date of Analysis: 11/02/2025 – 12/02/2025

Parameter Requirement: Physical, Chemical and Microbiological Analysis

Physical & Chemical Test	Specification	Results
Color	Light to medium, yellow to tan, trace hazy to slightly hazy	Passed
pH	7.30 ± 0.1	Passed
Hardness	Semi-solid with a firm, smooth, & bounces surface	Passed
Pour Depth/Thickness	4.00-4.50mm	Passed

Microbiological Test	Antimicrobial Agent	Zone Diameter Breakpoints Acceptable Range (mm)	MIC Breakpoints Acceptable Range
<i>Escherichia coli</i> ATCC® 25922	Penicillin 10 units	26-37	
	Gentamicin 10 µg	19-27	
	Erythromycin 15 µg	22-30	
	Clindamycin 2 µg	24-30	
	Trimethoprim-sulfamethoxazole 1.25/23.75 µg	24-32	
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Ceftazidime 30 µg	22-29	
	Amikacin 30 µg	20-26	
	Cefepime 30 µg	25-31	
	Imipenem 10 µg	20-28	
	Ciprofloxacin 5 µg	25-33	
	Gentamicin 10 µg	17-23	
	Colistin	-	0.25-2
<i>Enterococcus faecalis</i> ATCC® 29212	Gentamicin 10 µg	19-26	
	Amikacin 30 µg	19-26	
	Ampicillin 10 µg	15-22	
	Ceftazidime 30 µg	25-32	
	Penicillin	-	1-4
<i>Staphylococcus aureus</i> ATCC® 25923	Ampicillin 10 µg	15-21	
	Vancomycin 30 µg	10-16	
	Trimethoprim-sulfamethoxazole 1.25/23.75 µg	≥20	
	Nitrofurantoin 300 µg	18-24	
<i>Staphylococcus aureus</i> ATCC® 29213	Vancomycin	-	0.5-2
Uninoculated plate	No growth	-	-

Sample statement:

1. Culture media were prepared according to procedure SP-PD-BMDX-02 Production Process Control and SP-PD-BMDX-11 Product Specification.
2. Culture media were kept at temperature conditions 2-8°C.
3. QC Organisms were prepared and followed according to procedure SP-QC-BMDX-07 Quality Control Testing and SP-PD-BMDX-11 Product Specification.
4. * Pink Precipitation/Decoloration observed, QC still proceed to check the media performance.

Result:

The information given is believed to be correct, all results reported in this certificate relate only to the product in this certificate of analysis according to SP-PD-BMDX-11 Product Specification.

Sign by:

Fairuz Ayuni Binti Abdul Halim

QC Executive

BSc. in Microbiology with Honours
Universiti Putra Malaysia



Biomed MDX Sdn Bhd
8, Jalan IAN 3, Industri Angkasa Nuri,
76100 Durian Tunggal, Melaka, Malaysia

Date: 12/02/2025

