

Certificate of Analysis

Certificate No: BMDX/COA/P/056/2025

Product Name: Mueller Hinton II with 5% Sheep Blood

Product Code: FP90M2021

Lot No: 250704002

Date of Analysis: 09/07/2025 – 10/07/2025

Parameter Requirement: Physical, Chemical and Microbiological Analysis

Physical & Chemical Test	Specification	Results
Color	Red	Passed
pH	7.30 ± 0.1	Passed
Hardness	Semi-solid with a firm, smooth, & bounces surface	Passed
Pour Depth/Thickness	4.00-4.50mm	Passed

Microbiological Test	Antimicrobial Agent	Zone Diameter Breakpoints Acceptable Range (mm)	MIC Breakpoints Acceptable Range (µg/mL)
<i>Streptococcus pneumoniae</i> <i>ATCC® 49619</i>	Clindamycin 2 µg	19-25	
	Erythromycin 15 µg	25-30	
	Trimethoprim-sulfamethoxazole 75 µg	20-28	
	Penicillin		0.25-1
Uninoculated plate	No growth		

Sample statement:

- Culture media were prepared according to procedure SP-PD-BMDX-02 Production Process Control and SP-PD-BMDX-11 Product Specification.
- Culture media were control at temperature conditions 2-8°C.
- QC Organisms were prepared and followed according to procedure SP-QC-BMDX-07 Quality Control Testing and SP-PD-BMDX-11 Product Specification.

Result:

The information given is believed to be correct, all results reported in this certificate relate only to the product in this certificate of analysis according to SP-PD-BMDX-11 Product Specification.

Sign by:



Fairuz Ayuni Binti Abdul Halim

QC Executive

BSc. in Microbiology with Honours

Universiti Putra Malaysia

Date: 11/07/2025



Biomed MDX Sdn Bhd
8, Jalan IAN 3, Industri Angkasa Nuri,
76100 Durian Tunggal, Melaka, Malaysia