

Certificate of Analysis

Certificate No: BMDX/COA/P/018/2025

Product Name: Sabouraud Dextrose Agar

Product Code: FP90S1002

Lot No: 250318001

Date of Analysis: 18/03/2025 – 21/03/2025

Parameter Requirement: Physical, Chemical and Microbiological Analysis

Physical & Chemical Test	Specification	Results
Color	Light straw to straw-colored gel.	Passed
pH	5.60 ± 0.2	Passed
Hardness	Semi-solid with a firm, smooth, & bounces surface	Passed
Pour Depth/Thickness	4.00-4.50mm	Passed

Microbiological Test	Specification	Results
Candida albicans ATCC® 60193	Growth. Straw colonies	Passed
Trichophyton mentagrophytes ATCC® 9533	Growth. Pale straw-colored colonies.	Passed
Uninoculated plate	No growth	Passed

Sample statement:

1. Culture media were prepared according to procedure SP-PD-BMDX-02 Production Process Control and SP-PD-BMDX-11 Product Specification.
2. Culture media were kept at temperature conditions 2-8°C.
3. QC Organisms were prepared and followed according to procedure SP-QC-BMDX-07 Quality Control Testing and SP-PD-BMDX-11 Product Specification.

Result:

The information given is believed to be correct, all results reported in this certificate relate only to the product in this certificate of analysis according to SP-PD-BMDX-11 Product Specification.

Sign by:



Fairuz Ayuni Binti Abdul Halim

QC Executive

BSc. in Microbiology with Honours

Universiti Putra Malaysia

Date: 21/03/2025

