

Certificate of Analysis

Certificate No: BMDX/COA/P/004/2025

Product Name: CLED Agar / MacConkey III Agar

Product Code: SP90C1012

Lot No: 250206002

Date of Analysis: 11/02/2025 – 12/02/2025

Parameter Requirement: Physical, Chemical and Microbiological Analysis

Physical & Chemical Test	Specification	Results
Color	CLED Agar: Blue/Green-colored gel	Passed
	MacConkey III Agar: Dark red coloured gel	Passed
pH	CLED Agar: 7.30 ± 0.2	Passed
	MacConkey III Agar: 7.10 ± 0.2	Passed
Hardness	Semi-solid with a firm, smooth, & bounces surface	Passed
Pour Depth/Thickness	CLED Agar: 3.00mm to 3.50mm	Passed
	MacConkey III Agar: 3.00mm to 3.50mm	Passed

Microbiological Test	Specification	Results
CLED Agar:		
<i>E. coli</i> ATCC® 25922	Growth; colonies yellow, medium yellow	Passed
<i>P. mirabilis</i> ATCC® 12453	Growth; colonies blue, medium blue-green to blue	Passed
<i>S. aureus</i> ATCC® 25923	Growth; colonies small, yellow; medium yellow	Passed
<i>P. vulgaris</i> ATCC® 8427	Growth; colonies colorless to blue; swarming inhibited; slight spreading acceptable	Passed
Uninoculated plate	No growth	Passed
MacConkey III Agar:		
<i>E. coli</i> ATCC® 25922	Pink to red growth	Passed
<i>P. mirabilis</i> ATCC® 12453	Growth Colorless Inhibition of swarming	Passed
<i>S. Typhimurium</i> ATCC® 14028	Growth Colourless	Passed
<i>S. aureus</i> ATCC® 25923	No growth	Passed
<i>E. faecalis</i> ATCC® 29212	No growth	Passed
Uninoculated plate	No growth	Passed

Sample statement:

1. Culture media were prepared according to procedure SP-PD-BMDX-02 Production Process Control and SP-PD-BMDX-11 Product Specification.
2. Culture media were kept at temperature conditions 2-8°C.
3. QC Organisms were prepared and followed according to procedure SP-QC-BMDX-07 Quality Control Testing and SP-PD-BMDX-11 Product Specification.



Result:
The information given is believed to be correct, all results reported in this certificate relate only to the product in this certificate of analysis according to SP-PD-BMDX-11 Product Specification.

Sign by:




Fairuz Ayuni Binti Abdul Halim

QC Executive

BSc. in Microbiology with Honours
Universiti Putra Malaysia

Date: 17/02/2025



Biomed MDX Sdn Bhd
8, Jalan IAN 3, Industri Angkasa Nuri,
76100 Durian Tunggal, Melaka, Malaysia